



CALIBRATION CERTIFICATE

DAFCHGX002140002

1 3
Page 1 of 3 Pages





DIRECTIONS OF CALIBRATION

DAFCHGX002140002

Certificate No.

2 3

Page 2 of 3 Pages

1. The results is ONLR " i 15





Result of Calibration

DAFCHGX002140002

3 , 3

Report No.

Page 3 of 3 Pages

1 (Appearance & Interaction)

(Normal)

2 (Detection indication error):

(Conclusion)

(Pass/Fail)

	mg/L)	mg/L)	%)	Limit
Detection value	10	10.15	1.5	±10%
	50	50.81	1.6	± 10%
	80	78.58	-1.8	± 10%

P

P

P

3 (Repeatability of detection):

(Conclusion)

(Pass/Fail)

Standard value	Repeatability	Limit
mg/L	%	%
50	1.5	3

P

Notes

(Directions of measurement uncertainty in the calibration)

1. JJF1059.1-2012

(Reference Document JJF1059.1-2012 Evaluation and Expression of Uncertainty in Measurement)

2. (The expanded uncertainty of the measurement) $rel=3.2\% \quad =2$

(The below is blank)

